



General Information

FLEXALLOY® 35022RV-45NT CBL B8 NP

Teknor Apex Company - Flexible Polyvinyl Chloride

Wednesday, August 28, 2019

Seneral		
Material Status	Commercial: Active	
Availability	 Africa & Middle East Asia Pacific	Europe Latin America
Features	BPA FreeE-beam SterilizableEthylene Oxide Sterilizable	 Latex Free No Animal Derived Components Non-Phthalate Plasticizer Non-Phthalate - Citrate Radiation Sterilizable Plasticizer
Agency Ratings	• EC 1907/2006 (REACH)	OEHHA (California) Proposition 65
RoHS Compliance	RoHS Compliant	
Forms	Pellets	
Processing Method	Extrusion	Injection Molding

ASTM & ISO Properties ¹				
Physical	Nominal Value	Unit	Test Method	
Density / Specific Gravity	1.15		ASTM D792	
Mechanical	Nominal Value	Unit	Test Method	
Tensile Strength			ASTM D638	
50% Strain	210	psi		
100% Strain	360	psi		
300% Strain	900	psi		
Tensile Strength (Break)	1200	psi	ASTM D638	
Tensile Elongation (Break)	430	%	ASTM D638	
Hardness	Nominal Value	Unit	Test Method	
Durometer Hardness			ASTM D2240	
Shore A, 15 sec, Operating Stand	40			
Shore A, 15 sec, Hand Held	45			

FLEXALLOY® 35022RV-45NT CbL DO INF

Teknor Apex Company - Flexible Polyvinyl Chloride

Additional Information

EXAMPLE EXTRUSION APPLICATIONS Blood Pressure Tubing Blood Transport & Delivery Bubble Tubing Cable Jacketing Catheters Dialysis Tubing Drainage Tubing Endotracheal Airway Tubing Enteral Feeding Systems Infusion Tubing Oxygen Tubing Peristaltic Pump Tubing Suction Tubing

EXAMPLE INJECTION MOLDING APPLICATIONS

Adaptors **Blood Pressure Bulbs** Canulae Caps Connectors **Drip/Sight Chambers** Ear Protection Endotracheal Airway Cuffs Luers Mouthpieces Oxygen Masks Personal Safety Goggles **Resuscitation Bags** Suction Bulbs Syringe Bulbs Valves

Data Sheet Applies to Color(s) Listed Below: Clear Blue B8 NP (1136616)

Processing Information

 Extrusion
 Nominal Value
 Unit

 Melt Temperature
 335 to 350 °F

<u>+135-3858-6433</u> +188-1699-6168

Notes

¹ Typical properties: these are not to be construed as specifications.

Revision Date: 8/9/2019

The information and recommendations contained in this bulletin are, to the best of our knowledge, accurate and reliable but no guarantee of their accuracy is made. All products are sold upon condition that purchasers shall make their own tests to determine the suitability of such products for their particular purposes and uses and purchasers assume all risks and liability for the results of use of the products, including use in accordance with seller's recommendations. Nothing in this bulletin constitutes permission or a recommendation to practice or use any invention covered by any patent owned by this company or by others. There is no warranty of merchantability and there are no other warranties for the products described.