



FLEXALLOY® 35022RV-55NT CBL B8 NP

Teknor Apex Company - Flexible Polyvinyl Chloride

Wednesday, August 28, 2019

	General Info	ormation	
General			
Material Status	Commercial: Active		
Availability	 Africa & Middle East Asia Pacific	Europe Latin America North America	
Features	BPA FreeE-beam SterilizableEthylene Oxide Sterilizable	 Latex Free No Animal Derived Components Non-Phthalate Plasticizer Radiation (Gamma) Resistant Radiation Sterilizable 	
Agency Ratings	• EC 1907/2006 (REACH)	 OEHHA (California) Proposition 65 	
RoHS Compliance	 RoHS Compliant 		
Forms	• Pellets		
Processing Method	• Extrusion	Injection Molding	
	ASTM & ISO F	Properties ¹	
Physical		Nominal Value Un	it Test Method
Density / Specific Gravity		1.19	ASTM D792
Mechanical		Nominal Value Un	it Test Method
Tensile Strength			ASTM D638
50% Strain		305 psi	
100% Strain		520 psi	
300% Strain		1250 psi	
Tensile Strength (Break)		1700 psi	ASTM D638
Tensile Elongation (Break)		410 %	ASTM D638
Hardness		Nominal Value Un	it Test Method
Durometer Hardness			ASTM D2240
Shore A, 15 sec, Operating Stand		50	
Shore A, 15 sec, Hand Held		55	

+135-3858-6433 +188-1699-6168 +852-695-75415

FLEXALLOY® 35022RV-55NT CLEDO INF

Teknor Apex Company - Flexible Polyvinyl Chloride

Additional Information

EXAMPLE EXTRUSION APPLICATIONS

Blood Pressure Tubing

Blood Transport & Delivery

Bubble Tubing

Cable Jacketing

Catheters

Dialysis Tubing

Drainage Tubing

Endotracheal Airway Tubing

Enteral Feeding Systems

Infusion Tubing

Oxygen Tubing

Peristaltic Pump Tubing

Suction Tubing

EXAMPLE INJECTION MOLDING APPLICATIONS

Adaptors

Blood Pressure Bulbs

Canulae

Caps

Connectors

Drip/Sight Chambers

Ear Protection

Endotracheal Airway Cuffs

Luers

Mouthpieces

Oxygen Masks

Personal Safety Goggles

Resuscitation Bags

Suction Bulbs

Syringe Bulbs

Valves

Data Sheet Applies to Color(s) Listed Below:

Clear Blue B8 NP (1136614)

Processing Information Nominal Value Unit 345 to 360 °F

Notes

Extrusion

Melt Temperature

¹ Typical properties: these are not to be construed as specifications.

Revision Date: 8/9/2019