

# Geon™ Vinyl Flexible HC28500 Flexible Polyvinyl Chloride

# **Key Characteristics**

#### Product Description

Geon HC28500 is a 85 Shore A non-phthalate flexible vinyl compound made from FDA sanctioned ingredients. HC28500 can be used for extrusion and injection molding healthcare applications.

be used for extrusion and my	ection moiding healthcare applications.	
General		
Material Status	Commercial: Active	
Regional Availability	<ul> <li>Africa &amp; Middle East</li> <li>Europe</li> <li>Asia Pacific</li> <li>Latin America</li> <li>North America</li> </ul>	erica
Features	<ul> <li>BPA Free</li> <li>Good Processability</li> <li>Ethylene Oxide Sterilizable</li> <li>High Clarity</li> <li>Non-Phth</li> </ul>	alate Plasticizer
Uses	<ul> <li>Connectors</li> <li>Disposable Hospital Goods</li> <li>Hospital Goods</li> <li>Medical Devices</li> <li>Medical/Healthcare Applications</li> <li>Personal Care</li> <li>Tubing</li> </ul>	
Agency Ratings	FDA Unspecified Rating	
Appearance	Clear/Transparent	
Forms	• Pellets	
Processing Method	Extrusion     Injection Molding	

### Technical Properties 1

Physical	Typical Value (English)	Typical Value (SI)	Test Method
Density / Specific Gravity	1.26	1.26	ASTM D792
Mechanical	Typical Value (English)	Typical Value (SI)	Test Method
Tensile Strength <sup>2</sup> (100% Strain)	1700 psi	11.7 MPa	ASTM D638
Tensile Strength <sup>2</sup> (Break)	2900 psi	20.0 MPa	ASTM D638
Tensile Elongation <sup>2</sup> (Break)	370 %	370 %	ASTM D638
Hardness	Typical Value (English)	Typical Value (SI)	Test Method
Durometer Hardness (Shore A, 15 sec)	85	85	ASTM D2240
Thermal	Typical Value (English)	Typical Value (SI)	Test Method
Brittleness Temperature	-2.00 °F	-18.9 °C	ASTM D746
A 1 100 1 1 5 10			

## Additional Information

NOTE: All ingredients used in the formulation of this compound are listed in the United States FDA Code of Federal Regulations, Title 21. The only restriction is that the plasticizer is "For foods of high water content only". This statement is only an overview of the FDA listings for this product. It is the customer's responsibility to ensure that the compound is suitable for its intended application.

This product is not recommended for long-term storage of blood or blood components. Please contact PolyOne for more information.

#### Notes

Rev: 2016-05-19 Page: 1 of 2

<sup>&</sup>lt;sup>1</sup> Typical values are not to be construed as specifications.

<sup>&</sup>lt;sup>2</sup> Type IV, 20 in/min (510 mm/min)